

## Air Ionizer Verification Record

Ionizer Verification Sequence Number: 08-144

WORKING STANDARD USED						
Asset/ISO #:	Manufacturer:	Model:	Serial No.	Calibration Date:	Calibration Due:	Calibration By:
25171	ION	775	6779	8-20-08	8-20-09	JPL

AIR IONIZER INFORMATION						
Asset/ISO #:	Manufacturer:	Model:	Serial No.	Verification Date:	Verification Due:	Verification By:
29335 & 29336	ION	6442	8920 & 8924	10-13-08	2-25-09	JPL
Inspector:	Location:	Owner:	Fail: Y/N ?	Cleaned: Y/N ?	Adjusted: Y/N ?	Prior Sequence#
Hinh Do	170/213	Kari Lewis	N	N	N	08-113 & 114

VERIFICATION DATA						
HBM Sensitivity Level: <u>50</u> ✓ (from Table 1)						
Fan controller setting: <u>High</u> (High, Low, NA)						
Distance of ionizer from the charge plate: <u>24"</u>						
Ionizer Float Potential Tolerance $\pm$ <u>50</u> Vdc. (from Table 1)						
Measured Float Potential values recorded below.						
1	2	3	4	5	Comments:	
<u>10</u> Vdc.	<u>20</u> Vdc.	<u>10</u> Vdc.	<u>10</u> Vdc.	<u>10</u> Vdc.		
Ionizer Discharge Voltage Range: $\pm$ 1000 Vdc to $< \pm$ <u>50</u> Vdc (from Table 1)						
Ionizer Discharge Time Tolerance: <u>&lt;20</u> seconds. (from Table 1)						
Measured Discharge Time in second(s) and recorded values below.						
1 (+1000 to +Vdc)	2 (+1000 to +Vdc)	3 (+1000 to +Vdc)	4 (+1000 to +Vdc)	5 (+1000 to +Vdc)	Comments:	
<u>7.7</u> sec	<u>9.2</u> sec	<u>9.2</u> sec	<u>9.0</u> sec	<u>9.4</u> sec		
1 (-1000 to -Vdc)	2 (-1000 to -Vdc)	3 (-1000 to -Vdc)	4 (-1000 to -Vdc)	5 (-1000 to -Vdc)	Comments:	
<u>10.4</u> sec	<u>10.7</u> sec	<u>10.1</u> sec	<u>10.0</u> sec	<u>10.4</u> sec		

**Record** any corrective action required to restored ionizer operation (cleaning, adjustment, replacement, etc.)

If Ionizer was replaced, indicate below the identification of replacement.

Asset/ISO #: \_\_\_\_\_ Manufacturer: \_\_\_\_\_ Model: \_\_\_\_\_ Serial No.: \_\_\_\_\_

Sequence number for verification of replacement ionizer: \_\_\_\_\_

**Record** inspection schedule and rational for that schedule.